

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000346

Service date: 18.02.2026

Valid until: 18.02.2027

Result: **Approved**

Equipment details

Customer: Otech AS
Producer: Novasub
Serial no: 371129013
Customer tag: Diver 2

Location: Transporter
Model: Pneumo 0-60M
Build year: 2024
Service interval: 12 months

Calibration Instrument

Input basis: msw

Range: 0 + 60 m

Full scale: 60 msw

Accuracy (%FS): 0.25

Bar → msw factor: 1

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

| # | Instrument (msw) | Reference (bar) | Dev (msw) | %FS | Result |
|---|------------------|-----------------|-----------|--------|--------|
| 1 | 5.0 | 5.01 | 0.010 | 0.0167 | PASS |
| 2 | 10.0 | 10.01 | 0.010 | 0.0167 | PASS |
| 3 | 20.0 | 20.026 | 0.026 | 0.0433 | PASS |
| 4 | 30.0 | 30.023 | 0.023 | 0.0383 | PASS |
| 5 | 40.0 | 40.033 | 0.033 | 0.0550 | PASS |
| 6 | 50.0 | 50.03 | 0.030 | 0.0500 | PASS |

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

