

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000376

Service date: 21.02.2026

Valid until: 21.02.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: -
Serial no: 005 chamber
Customer tag: Chamber Panel

Location: Mjø
Model: Gauge
Build year: -
Service interval: 12 months

Calibration Instrument

Input basis: bar

Range: 100 bar

Full scale: 100 bar

Accuracy (%FS): 1.6

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (bar)	Reference (bar)	Dev (bar)	%FS	Result
1	20.0	20.19	0.190	0.1900	PASS
2	40.0	40.16	0.160	0.1600	PASS
3	60.0	60.31	0.310	0.3100	PASS
4	80.0	80.38	0.380	0.3800	PASS
5	100.0	100.81	0.810	0.8100	PASS
6	100.0	100.81	0.810	0.8100	PASS

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

