

# INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000528

Service date: 24.02.2026

Valid until: 24.02.2027

Result: **Approved**

## Equipment details

Customer: Sjøentreprenøren AS  
Producer: Gauge  
Serial no: 002 Panel  
Customer tag: Bank 3 inlet

Location: Mjø  
Model: Gauge HP  
Build year: —  
Service interval: 12 months

## Calibration Instrument

Input basis: bar

Range: 250 bar

Full scale: 250 bar

Accuracy (%FS): 1.6

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

## Verification results (As-Found)

#	Instrument (bar)	Reference (bar)	Dev (bar)	%FS	Result
1	50.0	48.5	1.500	0.6000	PASS
2	100.0	98.3	1.700	0.6800	PASS
3	150.0	149.5	0.500	0.2000	PASS
4	200.0	197.1	2.900	1.1600	PASS
5	250.0	247.0	3.000	1.2000	PASS
6	0.0	0.0	0.000	0.0000	PASS

## Technician

Serviced by: Daniel Kristensen

Signature

*Daniel Kristensen*

