

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000202

Service date: 21.01.2026

Valid until: 21.01.2027

Result: **Approved**

Equipment details

Customer: Samba Marin AS

Producer: AXSUB

Serial no: 005-1510

Customer tag: AXSub smart camera depth sensor

Location: Fox Inspector

Model: Depthsensor

Build year: 2023

Service interval: 12 months

Calibration Instrument

Input basis: msw

Accuracy (%FS): 1

Range: 0 + 50 m

Bar → msw factor: 1

Full scale: 50 msw

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	5.005	0.005	0.0100	PASS
2	10.0	10.09	0.090	0.1800	PASS
3	20.0	20.09	0.090	0.1800	PASS
4	30.0	30.2	0.200	0.4000	PASS
5	40.0	40.3	0.300	0.6000	PASS
6	50.0	50.11	0.110	0.2200	PASS

Technician

Serviced by: Daniel Kristensen

Signature

Electronically signed by DS Bergen AS

