

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000766

Service date: 24.06.2026

Valid until: 24.06.2027

Result: **Approved**

Equipment details

Customer: Frøy AS
Producer: Novasub
Serial no: 198250009
Customer tag: Diver 1

Location: 10Ft henger
Model: Pneumo 0-60M
Build year: 2026
Service interval: 12 months

Calibration Instrument

Input basis: msw
Accuracy (%FS): 0.25

Range: 0 + 70 m
Bar → msw factor: 10

Full scale: 70 msw

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS
Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	0.5001	0.001	0.0014	PASS
2	10.0	1.0024	0.024	0.0343	PASS
3	20.0	2.0028	0.028	0.0400	PASS
4	30.0	3.0014	0.014	0.0200	PASS
5	50.0	5.0019	0.019	0.0271	PASS

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

