

# HYPERBARIC CHAMBER SERVICE CERTIFICATE

DSC-2026-000185

Service date: 16.12.2025

Valid to: 16.12.2026

Result: **Approved**

## Equipment details

Customer: Sjøentreprenøren AS  
Producer: Draeger  
Serial no: 1060  
Service interval: 12 months

Location: Kammerkontainer  
Model: Chamber  
Build year: 1977

## Pressure

Pressure reducing valve	20.0	Pass
Main chamber overpressure valve	5.5	Pass
Entry lock overpressure valve	—	-
Pipe maximum working pressure	200.0	Pass

Pressure test  
Pressure: — bar  
Time: — min  
Descent: — min  
Ascent: — min

Hydro  
Last hydrostatic test:  
01.01.2020  
Next hydrostatic test:  
01.01.2030  
Hydrotested: No  
Testrange: —

## Checklist

General condition	Pass	Communication – main chamber	Pass
Entry lock inlet valve	Pass	Communication – Entry lock	Pass
Entry lock outlet valve	Pass	Gasket – medical lock	Pass
Main chamber inlet valve	Pass	Gasket – Entry lock door	Pass
Main chamber outlet valve	Pass	Gasket – main chamber door	Pass
Depth alarm – audible signal	-	Window/viewport New 2020	Pass
Depth gauge	Pass	Valves & Piping	Pass
High pressure gauge (HP)	Pass	Lightening	Pass
Low pressure gauge (LP)	Pass	Cameras	-
BIBS / O <sub>2</sub> mask	Pass	Surface treatment inside/outside	Pass
LP bank	-	Equalizer	Pass
HP Bank	Pass	Drainage	Pass
Next hydro 2033			
O <sub>2</sub> needle valves internal / external	Pass		

## Notes

Chamber total rehab 2024 by DS Bergen.

## Technician

Serviced by: Daniel Kristensen

Signature



# HYPERBARIC CHAMBER SERVICE CERTIFICATE

DSC-2026-000185

Service date: 16.12.2025

Valid to: 16.12.2026

Result: **Approved**

Daniel K. Gibson