

HYPERBARIC CHAMBER SERVICE CERTIFICATE

DSC-2026-000615

Service date: 08.04.2026

Valid to: 08.04.2027

Result: **Approved**

Equipment details

Customer: Valen marina AS
Producer: Draeger
Serial no: 1244
Service interval: 12 months

Location: Chamber Container
Model: Chamber
Build year: 1982

Pressure

Pressure reducing valve	—	Pass
Main chamber overpressure valve	5.5	Pass
Entry lock overpressure valve	—	-
Pipe maximum working pressure	300.0	Pass

Pressure test
Pressure: — bar
Time: — min
Descent: — min
Ascent: 0 min

Hydro
Last hydrostatic test:
30.08.2023
Next hydrostatic test:
30.08.2033
Hydrotested: No
Testrange: —

Checklist

General condition	Pass	Communication – main chamber	Pass
Entry lock inlet valve	Pass	Communication – Entry lock	Pass
Entry lock outlet valve	Pass	Gasket – medical lock	Pass
Main chamber inlet valve	Pass	Gasket – Entry lock door	Pass
Main chamber outlet valve	Pass	Gasket – main chamber door	Pass
Depth alarm – audible signal DE-OX	Pass	Window/viewport 06/2023	Pass
Depth gauge	Pass	Valves & Piping	Pass
High pressure gauge (HP)	Pass	Lightening	Pass
Low pressure gauge (LP)	Pass	Cameras	Pass
BIBS / O ₂ mask see bibs certificate	Pass	Surface treatment inside/outside	Pass
LP bank	-	Equalizer	Pass
HP Bank Composite 8x50L 200Bar	Pass	Drainage	Pass
O ₂ needle valves internal / external	Pass		

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

