

ANALYZER CALIBRATION

Certificate no: DSC-2026-000621

Service date: 08.04.2026

Valid until: 08.04.2027

Result: **Approved**

Equipment details

Customer: Valen marina AS

Producer: Deox

Serial no: GA01

Service interval: 12 months

Location: Chamber Container

Model: Multigas

Build year: -

NOTE

We confirm that Above analyzer has been Tested and calibrated with 100% oxygen and 20,9% air. Oxygen sensor has been replaced before calibration. 100% o2 - Adjusted to 100% 20,9% air - Adjusted to 20,9%

Technician

Serviced by: Daniel Kristensen

Signature

