

DIVATOR MKII SET

Certificate no: DSC-2026-000679

Service date: 28.04.2026

Valid until: 28.04.2027

Result: **Approved**

Equipment details

Customer: Havforskningsinstituttet

Producer: Interspiro

Serial no: AGA 09 & 9215346

Service interval: 12 months

Location: Flødevigen

Model: Divator MKII

Build year: -

NOTE

Tested with Labtec PRO

This is a certificate confirmation, the documentation has a separate file.

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen



test report

test report: 467
 device no.: 1639500
 type of Yearly test
 test system: PRO

3102.24.325

workshop: DS BERGEN AS		person: Daniel Kristensen			
street: Storebotn 55G		phone: 91649162			
place: 5309 Kleppestø		email: Daniel@dsbergen.no			
customer: Havforskningsinstituttet		person: Børge Alfstad			
street: Nordnesgaten 50		phone: 93201295			
place: 5005 Bergen		email: boerge.alfstad@hi.no			
device data:	deadlines:	device-info:	service:		
device no: 1639500	1 y test 28.04.2027 1 y srv 28.04.2027 6 y srv 13.03.2031		- 1 year service - cleaning and disinfection		
type: Interspiro Divator PP					
class: mask PP					
location: Flødevigen					
Complete Divator mask and regulator with Apeks XTX 1.stage. Apeks 2.stage as bailout regulator					
tests/measurements					
	unit	min	max	actual	result
visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...)			0	0	OK
mask fitted approx. 6 sec head expansion Mask volume 0,2 l				0	OK
leakage test w/ open exhalation valve 0,25 min at p=-10 mbar	mbar		7,5	2,4	OK

test executed: 28.04.2026 12:39
 by: Daniel Kristensen
 test signed: 28.04.2026 12:39
 by: Daniel Kristensen
 LabNet V: 2025.12.2.0
 next test on: 28.04.2027

result:
 device ready for use

test report

test report: 467
 device no.: 9215346
 type of Yearly test
 test system: PRO 3102.24.325

workshop: DS BERGEN AS		person: Daniel Kristensen			
street: Storebotn 55G		phone: 91649162			
place: 5309 Kleppestø		fax:			
		email: Daniel@dsbergen.no			
customer: Havforskningsinstituttet		person: Børge Alfstad			
street: Nordnesgaten 50		phone: 93201295			
place: 5005 Bergen		fax:			
		email: boerge.alfstad@hi.no			
device data:	deadlines:	device-info:	service:		
device no: 9215346	1 y test 28.04.2027 1 y srv 28.04.2027 6 y srv 13.03.2031		1 year service cleaning and disinfection		
type: Interspiro BV Divator NP					
class: demand valve NP					
location: Flødevigen					
Complete Dlvator mask and regulator with Apeks XTX 1.stage. Apesk 2.stage as bailout regulator					
tests/measurements	unit	min	max	actual	result
visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...)			0	0	OK
demand valve opening pressure at 10 l/min suction	mbar	-4	-0,5	-1,6	OK
inhalation resistance at 40 x 2,5 l/min, 5 hubs HP .>. bar	mbar	-8	-0,1	-4	OK
exhalation valve opening pressure at 10 l/min suction	mbar	2	5	2,3	OK
safety valve opening pressure	bar	12	20	14,5	OK
pressure release of test equipment (HP, MP and LP)				0	OK

test executed: 28.04.2026 12:39
 by: Daniel Kristensen
 test signed: 28.04.2026 12:39
 by: Daniel Kristensen

result:
 device ready for use

LabNet V: 2025.12.2.0

next test on: 28.04.2027