

OXYBOX SERVICE CERTIFICATE

DSC-2026-000165

Service date: 07.01.2026

Valid until: 07.01.2027

Result: **Approved**

Equipment details

Customer: IMC Diving AS
Producer: CBM
Serial no: 719
Service interval: 12 months

Location: 719
Model: Oxybox
Build year: 2019

| Cylinder size | Hydro test | Next hydro test | Filled date | Oxygen cleaned |
|------------------------|-------------------|-------------------------|-------------|----------------|
| 5 | 30.06.2019 | 30.06.2029 | — | - |
| 1st stage SN 466831 | 2nd stage SN — | Cylinder SN EED151UT | | |

Service items

| | | | |
|--------------------------|------|--------------------------------------|------|
| General condition | Pass | Equipment leak test | Pass |
| O ₂ regulator | Pass | Oxygen analysis (-) | - |
| Gauge | Pass | Pocket mask | Pass |
| Hose | Pass | Freeflow mask | Pass |
| Comment: new 2019 | | Comment: Free flow mask Changed 2026 | |
| Equipment cleaned | Pass | Flow test regulator | Pass |

Technician

Serviced by: Daniel

Signature

Daniel K. Eriksen

