

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000380

Service date: 22.02.2026

Valid until: 22.02.2027

Result: **Approved**

## Equipment details

Customer: Sjøentreprenøren AS  
Producer: CBM  
Serial no: 080172917A  
Service interval: 12 months

Location: Mjø  
Model: Oxybox  
Build year: —

Cylinder size 5L	Hydro test 01.11.2017	Next hydro test 01.11.2027	Filled date 01.03.2024	Oxygen cleaned No
1st stage SN 210568081	2nd stage SN 20061214	Cylinder SN AIL137UT		

## Service items

General condition	Pass	Equipment leak test	Pass
O <sub>2</sub> regulator	Pass	Oxygen analysis (99.9)	Pass
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Comment: changed	
		Flow test regulator	Pass

## Technician

Serviced by: Daniel Kristensen

Signature

*Daniel Kristensen*

