

OXYBOX SERVICE CERTIFICATE

DSC-2026-000186

Service date: 16.12.2025

Valid until: 16.12.2026

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: CBM
Serial no: A060579554A
Service interval: 12 months

Location: Kammerkontainer
Model: Oxybox
Build year: —

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	19.11.2017	19.11.2027	—	-
1st stage SN	2nd stage SN	Cylinder SN		
—	A060579554A	AIL141UT		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Flow test regulator	Pass

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen