

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000204

Service date: 22.01.2026

Valid until: 22.01.2027

Result: **Approved**

## Equipment details

Customer: Kystarbeid AS  
Producer: SANOSUB  
Serial no: 064  
Service interval: 12 months

Location: MS Nitrox  
Model: Oxybox  
Build year: 2024

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.05.2024	01.05.2034	—	-
1st stage SN	2nd stage SN	Cylinder SN		
—	—	—		

## Service items

General condition	Pass	Equipment leak test	Pass
O <sub>2</sub> regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Comment: Freeflow mask changed	
Comment: Crimped new connector		Freeflow mask	-
Equipment cleaned	Pass	Flow test regulator	-

## Notes

Oxybox in Green tube.

## Technician

Serviced by: Daniel Kristensen

## Signature

Electronically signed by DS Bergen AS

