

OXYBOX SERVICE CERTIFICATE

DSC-2026-000600

Service date: 06.04.2026

Valid until: 06.04.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: CBM
Serial no: 1056
Service interval: 12 months

Location: Alu kontainer
Model: Oxybox
Build year: —

Cylinder size 5L	Hydro test 01.03.2025	Next hydro test 01.03.2035	Filled date 06.04.2026	Oxygen cleaned Yes (06.04.2026)
1st stage SN 0803833378A	2nd stage SN 20070808	Cylinder SN 25AG062 021UT		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (99.6)	Pass
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Flow test regulator	Pass

Notes

New cylinder

Technician

Serviced by: Daniel Kristensen

Signature

