

OXYBOX SERVICE CERTIFICATE

DSC-2026-000214

Service date: 22.01.2026

Valid until: 22.01.2027

Result: **Approved**

Equipment details

Customer: IMC Diving AS
Producer: CBM
Serial no: 737
Service interval: 12 months

Location: 737
Model: Oxybox
Build year: 2012

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.08.2024	24.08.2034	—	-
1st stage SN 407302	2nd stage SN —	Cylinder SN J63083		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Comment: Freeflow mask changed	
Comment: Changed		Freeflow mask	Pass
Equipment cleaned	Pass	Flow test regulator	Pass

Notes

Oxybox yellow case
regulator SN 407302
Cylinder SN J63083

Technician

Serviced by: Daniel Kristensen

Signature

