

OXYBOX SERVICE CERTIFICATE

DSC-2026-000712

Service date: 27.05.2026

Valid until: 27.05.2027

Result: **Approved**

Equipment details

Customer: Dypvest AS

Producer: SANOSUB

Serial no: 1083

Service interval: 12 months

Location: Båt

Model: Oxybox

Build year: 2021

Cylinder size
5L

Hydro test
01.06.2021

Next hydro test
01.06.2031

Filled date
27.05.2026

Oxygen cleaned
No

1st stage SN
041

2nd stage SN
487

Cylinder SN
L45216

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (99.9)	Pass
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Comment: Changed	
		Flow test regulator	Pass

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

