

OXYBOX SERVICE CERTIFICATE

DSC-2026-000341

Service date: 17.02.2026

Valid until: 17.02.2027

Result: **Approved**

Equipment details

Customer: Otech AS
Producer: -
Serial no: RT112383715
Service interval: 12 months

Location: 10Ft
Model: Oxybox Bag
Build year: 2014

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.03.2025	01.03.2035	-	-
1st stage SN RT112383715	2nd stage SN ODV112307880	Cylinder SN 25AG0o2 010UT		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Comment: New 2025	
		Flow test regulator	Pass

Notes

New cylinder 2026. O2 cleaned and filled up to 200Bar with 100% Oxygen

Technician

Serviced by: Daniel Kristensen

Signature

