

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000351

Service date: 18.02.2026

Valid until: 18.02.2027

Result: **Approved**

## Equipment details

Customer: Otech AS  
Producer: CBM  
Serial no: 2072  
Service interval: 12 months

Location: Transporter  
Model: Oxybox  
Build year: 2015

Cylinder size 5L	Hydro test 01.03.2025	Next hydro test 01.03.2035	Filled date -	Oxygen cleaned -
1st stage SN 15-06/339	2nd stage SN -	Cylinder SN 25AG062 013UT		

## Service items

General condition	Pass	Equipment leak test	Pass
O <sub>2</sub> regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Flow test regulator	Pass

## Notes

New cylinder 2026.

## Technician

Serviced by: Daniel Kristensen

Signature

*Daniel Kristensen*

