

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000166

Service date: 13.01.2026

Valid until: 13.01.2027

Result: **Approved**

Equipment details

Customer: Dypvest AS
Producer: AXSUB
Serial no: 002-0084
Customer tag: –

Location: Båt
Model: Depthsensor
Build year: –
Service interval: 12 months

Calibration Instrument

Input basis: msw
Accuracy (%FS): 1

Range: 0 + 70 m
Bar → msw factor: 1

Full scale: 70 msw

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS
Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	4.9501	0.0499	0.0713	PASS
2	10.0	9.9	0.100	0.1429	PASS
3	20.0	19.82	0.180	0.2571	PASS
4	30.0	29.77	0.230	0.3286	PASS
5	40.0	39.71	0.290	0.4143	PASS
6	50.0	49.73	0.270	0.3857	PASS

Technician

Serviced by: Daniel

Signature

Daniel K. Sørensen

