

HYPERBARIC CHAMBER SERVICE CERTIFICATE

DSC-2026-000185

Service date: 16.12.2025

Valid to: 16.12.2026

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: Draeger
Serial no: 1060
Service interval: 12 months

Location: Kammerkontainer
Model: Chamber
Build year: 1977

Pressure

Pressure reducing valve	20.0	Pass
Main chamber overpressure valve	5.5	Pass
Entry lock overpressure valve	—	-
Pipe maximum working pressure	200.0	Pass

Pressure test
Pressure: — bar
Time: — min
Descent: — min
Ascent: — min

Hydro
Last hydrostatic test:
01.01.2020
Next hydrostatic test:
01.01.2030
Hydrotested: No
Testrange: —

Checklist

General condition	Pass	Communication – main chamber	Pass
Entry lock inlet valve	Pass	Communication – Entry lock	Pass
Entry lock outlet valve	Pass	Gasket – medical lock	Pass
Main chamber inlet valve	Pass	Gasket – Entry lock door	Pass
Main chamber outlet valve	Pass	Gasket – main chamber door	Pass
Depth alarm – audible signal	-	Window/viewport New 2020	Pass
Depth gauge	Pass	Valves & Piping	Pass
High pressure gauge (HP)	Pass	Lightening	Pass
Low pressure gauge (LP)	Pass	Cameras	-
BIBS / O ₂ mask	Pass	Surface treatment inside/outside	Pass
LP bank	-	Equalizer	Pass
HP Bank	Pass	Drainage	Pass
Next hydro 2033			
O ₂ needle valves internal / external	Pass		

Notes

Chamber total rehab 2024 by DS Bergen.

Technician

Serviced by: Daniel Kristensen

Signature



HYPERBARIC CHAMBER SERVICE CERTIFICATE

DSC-2026-000185

Service date: 16.12.2025

Valid to: 16.12.2026

Result: **Approved**

Daniel K. Gibson