

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000428

Service date: 23.02.2026

Valid until: 23.02.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: Gauge
Serial no: Y06 Diver 2
Customer tag: –

Location: Yankee
Model: Pneumo
Build year: –
Service interval: 12 months

Calibration Instrument

Input basis: msw

Range: 0 + 70 m

Full scale: 70 msw

Accuracy (%FS): 0.25

Bar → msw factor: 9.94

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

| # | Instrument (msw) | Reference (bar) | Dev (msw) | %FS | Result |
|---|------------------|-----------------|-----------|--------|--------|
| 1 | 5.0 | 0.502 | 0.0101 | 0.0144 | PASS |
| 2 | 10.0 | 0.995 | 0.1097 | 0.1567 | PASS |
| 3 | 20.0 | 2.005 | 0.0703 | 0.1004 | PASS |
| 4 | 30.0 | 3.007 | 0.1104 | 0.1577 | PASS |
| 5 | 40.0 | 4.03 | 0.0582 | 0.0831 | PASS |
| 6 | 50.0 | 5.02 | 0.1012 | 0.1446 | PASS |

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

