

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000195

Service date: 10.12.2025

Valid until: 10.12.2026

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: Novasub
Serial no: 371129014
Customer tag: Diver 2

Location: Kammerkontainer
Model: Pneumo 0-60M
Build year: 2024
Service interval: 12 months

Calibration Instrument

Input basis: msw

Range: 0 + 60 m

Full scale: 60 msw

Accuracy (%FS): 0.25

Bar → msw factor: 1

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	4.96	0.040	0.0667	PASS
2	10.0	9.97	0.030	0.0500	PASS
3	20.0	19.99	0.010	0.0167	PASS
4	30.0	30.0	0.000	0.0000	PASS
5	40.0	40.02	0.020	0.0333	PASS
6	50.0	50.02	0.020	0.0333	PASS

Remarks

Mounted divepanel SN 2405 - Diver 2

Technician

Serviced by: Daniel Kristensen

Signature

Electronically signed by DS Bergen AS

