

# INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000619

Service date: 08.04.2026

Valid until: 08.04.2027

Result: **Approved**

## Equipment details

Customer: Valen marina AS

Producer: Gauge

Serial no: Entry Lock

Customer tag: Chamber Entry lock panel

Location: Chamber Container

Model: Pneumo

Build year: —

Service interval: 12 months

## Calibration Instrument

Input basis: msw

Accuracy (%FS): 0.25

Range: 0 + 60 m

Bar → msw factor: 10

Full scale: 60 msw

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +/-2°C Traceability to international standards

## Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	0.501	0.010	0.0167	PASS
2	10.0	1.012	0.120	0.2000	PASS
3	20.0	2.011	0.110	0.1833	PASS
4	30.0	3.0113	0.113	0.1883	PASS
5	50.0	5.0112	0.112	0.1867	PASS

## Technician

Serviced by: Daniel Kristensen

Signature

