

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000214

Service date: 22.01.2026

Valid until: 22.01.2027

Result: **Approved**

## Equipment details

Customer: IMC Diving AS  
Producer: CBM  
Serial no: 737  
Service interval: 12 months

Location: 737  
Model: Oxybox  
Build year: 2012

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.08.2024	24.08.2034	—	-
1st stage SN 407302	2nd stage SN —	Cylinder SN J63083		

## Service items

General condition	Pass	Equipment leak test	Pass
O <sub>2</sub> regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Comment: Freeflow mask changed	
Comment: Changed		Freeflow mask	Pass
Equipment cleaned	Pass	Flow test regulator	Pass

## Notes

Oxybox yellow case  
regulator SN 407302  
Cylinder SN J63083

## Technician

Serviced by: Daniel Kristensen

Signature

*Daniel Kristensen*

