

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000712

Service date: 27.05.2026

Valid until: 27.05.2027

Result: **Approved**

## Equipment details

Customer: Dypvest AS

Producer: SANOSUB

Serial no: 1083

Service interval: 12 months

Location: Båt

Model: Oxybox

Build year: 2021

|                     |                          |                               |                           |                      |
|---------------------|--------------------------|-------------------------------|---------------------------|----------------------|
| Cylinder size<br>5L | Hydro test<br>01.06.2021 | Next hydro test<br>01.06.2031 | Filled date<br>27.05.2026 | Oxygen cleaned<br>No |
| 1st stage SN<br>041 | 2nd stage SN<br>487      | Cylinder SN<br>L45216         |                           |                      |

## Service items

|                          |      |                        |      |
|--------------------------|------|------------------------|------|
| General condition        | Pass | Equipment leak test    | Pass |
| O <sub>2</sub> regulator | Pass | Oxygen analysis (99.9) | Pass |
| Gauge                    | Pass | Pocket mask            | Pass |
| Hose                     | Pass | Freeflow mask          | Pass |
| Equipment cleaned        | Pass | Comment: Changed       |      |
|                          |      | Flow test regulator    | Pass |

## Technician

Serviced by: Daniel Kristensen

Signature

*Daniel Kristensen*

